



Your Logo Here

**Vitrekt QT Insite 95x Test Report**

**Page 1**

**DUT Model: Your-Model**

**DUT Serial Number: A-Serial-Number**

**Tested By: tjw on 3/17/25,13:59:44**

**Instrument: 952i SN: 017909**

**Passed all tests**

Step #1 Hold: 60.0s PASSED

Step #2 ACW @ 220.0V/60.0Hz for 10.0s PASSED

Step #3 ACCAP @ 300.0V/60.0Hz for 10.0s PASSED

Step #4 DCW @ 400.0V for 11.0s PASSED

Step #5 DCIR @ 500.0V for 10.0s PASSED

Step #6 ACI for 10.0s PASSED

Step #7 DCI for 10.0s PASSED

Step #8 Hold: 60.0s PASSED

Step #9 LowOhms 2-wire for 5.0s PASSED

Step #10 GB @ 2.0A/60.0Hz for 10.0s PASSED

Step #11 Hold: 60.0s PASSED